



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/750,173	12/31/2003	Edward B. Zalenski	101896-719 (DEP-5165)	5023
21125 7590 03/02/2011 NUTTER MCCLENNEN & FISH LLP SEAPORT WEST 155 SEAPORT BOULEVARD BOSTON, MA 02210-2604				
EXAMINER				
ARAJ, MICHAEL J				
ART UNIT		PAPER NUMBER		
3775				
NOTIFICATION DATE		DELIVERY MODE		
03/02/2011		ELECTRONIC		

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

docket@nutter.com

UNITED STATES PATENT AND TRADEMARK OFFICE

---

BEFORE THE BOARD OF PATENT APPEALS  
AND INTERFERENCES

---

*Ex parte* EDWARD B. ZALENSKI, DALE W. FRANK,  
MICHAEL D. SORRENTI, and ALEXANDER GRINBERG

---

Appeal 2009-014783  
Application 10/750,173  
Technology Center 3700

---

Before TONI R. SCHEINER, DONALD E. ADAMS, and STEPHEN  
WALSH, *Administrative Patent Judges*.

WALSH, *Administrative Patent Judge*.

DECISION ON APPEAL<sup>1</sup>

This is an appeal under 35 U.S.C. § 134(a) involving claims to an implant device. The Patent Examiner rejected the claims on grounds of anticipation and obviousness. We have jurisdiction under 35 U.S.C. § 6(b). We reverse.

---

<sup>1</sup> The two-month time period for filing an appeal or commencing a civil action, as recited in 37 C.F.R. § 1.304, or for filing a request for rehearing, as recited in 37 C.F.R. § 41.52, begins to run from the “MAIL DATE” (paper delivery mode) or the “NOTIFICATION DATE” (electronic delivery mode) shown on the PTOL-90A cover letter attached to this decision.

## STATEMENT OF THE CASE

Claims 1-7, 9, 11-13 and 15 are on appeal. Claim 1 is representative and reads as follows:

1. An implant device, comprising:
  - (i) a frame having a trigger mechanism;
  - (ii) an outer sleeve mechanically coupled to the frame;
  - (iii) an inner shaft having a grabber for mechanically engaging an implant, the inner shaft slidably disposed along a major axis of the inner shaft within the outer sleeve, whereby actuation of the trigger extends the grabber from the outer sleeve to thereby release the implant; and
  - (iv) a retaining element for directing the grabber toward a closed position whereby the grabber is substantially contained within the outer sleeve when the trigger is released.

The Examiner rejected the claims as follows:

- claims 1-7, 9, 11, 12 and 15 under 35 U.S.C. § 102(e) as anticipated by Markworth; and
- claim 13 under 35 U.S.C. § 103(a) as unpatentable over Markworth and Lim.

## ANTICIPATION

### *The Issue*

Appellants contend that “Markworth . . . discloses a different device that is configured in a way that is the diametric opposite of relevant claim elements.” (App. Br. 4.)

The issue with respect to this rejection is whether the Markworth device is configured as Appellants’ device is configured.

### *Principles of Law*

“[U]nless a prior art reference discloses within the four corners of the document not only all of the limitations claimed but also all of the

limitations arranged or combined in the same way as recited in the claim, it cannot be said to prove prior invention of the thing claimed and, thus, cannot anticipate under 35 U.S.C. § 102.” *Net MoneyIN, Inc. v. Verisign, Inc.*, 545 F.3d 1359, 1371 (Fed. Cir. 2008).

#### *Analysis*

We agree with Appellants that the rejection did not provide evidence that Markworth described the device defined by the appealed claims. Appellants have persuasively explained that: (1) Markworth did not describe the outer sleeve mechanically coupled to a frame as defined in claim element (ii) (App. Br. 10-11); (2) Markworth did not describe an inner shaft having a grabber as defined in claim element (iii) (*id.* at 11-12); (3) Markworth did not describe an inner shaft slidably disposed within an outer sleeve as defined in claim element (iii) (*id.* at 12); and (4) rather than describing extension of a grabber by trigger actuation, Markworth described the opposite (*id.* at 13-14). Because Markworth did not describe all the elements arranged as claimed, we must reverse this rejection.

#### OBVIOUSNESS

The obviousness rejection of claim 13 relied on the same findings concerning Markworth’s device that we found insufficient as a description of the claimed device. The obviousness rejection did not account for all the differences between Markworth’s device and the claimed device, and we must therefore reverse it.

#### SUMMARY

We reverse the rejection of claims 1-7, 9, 11, 12 and 15 under 35 U.S.C. § 102(e) as anticipated by Markworth.

Appeal 2009-014783  
Application 10/750,173

We reverse the rejection of claim 13 under 35 U.S.C. § 103(a) as unpatentable over Markworth and Lim.

REVERSED

lp

UTTER MCCLENNEN & FISH LLP  
SEAPORT WEST  
155 SEAPORT BOULEVARD  
BOSTON MA 02210-2604